Doctor, Will My Surgical Hardware Set Off Metal Detector in the Airport?

Doutor, meus implantes irão ativar o detector de metais no aeroporto?

Igor Guedes Nogueira Reis[1] Beatriz Marinho Guimarães[1] Samuel Henrique Ferreira de Souza[1] Marco Antônio Percope de Andrade[2] Robinson Esteves Pires[2]

[1] Orthopedic Surgery Residency, Locomotor System Department, Hospital das Clínicas, Universidade Federal de Minas Gerais, Belo Horizonte, MG, Brazil
[2] Locomotor System Department, Hospital das Clínicas, Universidade Federal de Minas Gerais, Belo Horizonte, MG, Brazil

Address for correspondence Igor Guedes Nogueira Reis, Departamento do Aparelho Locomotor, Hospital das Clínicas, Universidade Federal de Minas Gerais. Endereço: Av. Prof. Alfredo Balena, 190–Santa Efigênia, Belo Horizonte, MG, 30130-100, Brazil (e-mail: dr.igorgnreis@gmail.com).

Rev Bras Ortop 2024;59(5):e672–e681.

Abstract

Objective Verify if routinely used metallic implants (stainless steel, aluminum alloy, cobalt-chromium-molybdenum, and titanium made) will be detected in an international airport of Brazil and generate helpful information to prevent patient inconvenience and to support the security regulatory agencies.

Methods An experimental, non-randomized, controlled, cross-over study was performed by recruiting two individuals, one male and one female, to pass through a standard airport metal detector with orthopedic implants attached to the body. Implants with different compositions, weight, and in various parts of the body were tested.

Results From all implants tested, there was no detection of implants for internal fixation, whether steel or titanium. The external fixator was detected and the only difference in composition is that the external fixator tested have aluminum alloy. All hip replacement implants tested were detected. Two knee replacement implants were tested, and both were made of cobalt-chromium-molybdenum, but with different specifications and only one of them was detected.

Conclusions In this study with ex-vivo orthopedic implants, we have found that osteosynthesis implants composed by Stainless Steel ISO 5832-1 did not trigger the airport walk-through metal detector. However, external fixator and total joint prostheses were more frequently detected.

Keywords
- airports
- internal fracture fixation
- metals
- patient education as topic
- prosthesis and implants

Work carried out at the Universidade Federal de Minas Gerais, Belo Horizonte, MG, Brazil.

received
October 9, 2022
accepted after revision
December 2, 2022

DOI https://doi.org/10.1055/s-0043-1771493.
ISSN 0102-3616.

© 2023. The Author(s).
This is an open access article published by Thieme under the terms of the Creative Commons Attribution 4.0 International License, permitting copying and reproduction so long as the original work is given appropriate credit (https://creativecommons.org/licenses/by/4.0/).

Thieme Revinter Publicações Ltda., Rua do Matoso 170, Rio de Janeiro, RJ, CEP 20270-135, Brazil

Resumo

Objetivo Verificar se implantes metálicos de uso rotineiro (aço inoxidável, liga de alumínio, cobalto-cromo-molibdênio e feitos de titânio) serão detectados em um aeroporto internacional do Brasil e gerar informações úteis para evitar transtornos ao paciente e apoiar os órgãos reguladores de segurança.

Métodos Estudo experimental, não randomizado, controlado e cruzado foi realizado recrutando dois indivíduos, um homem e uma mulher, para passar por um detector de metais padrão do aeroporto com implantes ortopédicos presos ao corpo. Foram testados implantes com diferentes composições, pesos e em várias partes do corpo.

Resultados De todos os implantes testados, não houve detecção de implantes para fixação interna, sejam de aço ou de titânio. O fixador externo foi detectado e a única diferença na composição é que o fixador externo testado é de liga de alumínio. Todos os implantes de artroplastia do quadril testados foram detectados. Dois implantes de artroplastia do joelho foram testados, e ambos eram feitos de cobalto-cromo-molibdênio, mas com especificações diferentes, e apenas um deles foi detectado.

Conclusões Neste estudo com implantes ortopédicos ex-vivo, verificamos que os implantes de osteossíntese compostos por aço inoxidável ISO 5832-1 não acionaram o arco detector de metais do aeroporto. No entanto, o fixador externo e as próteses articulares totais foram mais frequentemente detectadas.

Palavras-chave

- aeroportos
- educação de pacientes como assunto
- fixação interna de fraturas
- metais
- próteses e implantes

Introduction

An agreement on what are the recommendations for patients with implants in regard to air travel still lacks in current literature. However, due to the catastrophic incident of September 11th, 2001, data on this topic has arisen progressively.[1–4] This historical tragedy stimulated travel restrictions and sensitivity of airport metal detectors were increased, leading to inconvenience, such as thorough and prolonged body search reported by patients during national and international air travels.[5–9]

It is still a matter of controversy the factors that influence detection of some implants, but some hypothesis were generated by observational studies, surveys and experimental studies.[4] Metal composition, metal weight, amount of body mass, distance from metal to detector, and speed when crossing the detector are potential elements that can influence the metal detection. It is noteworthy that we did not find any study on this topic in the Brazilian literature, which confirms that the topic is worthy of investigation.

Therefore, we have performed this experimental study to foment the discussion in the Brazilian scientific community, investigate possible factors associated to orthopedic implant detection in airports and generate useful data for a better guidance to patients and to the air-security agencies.

Materials and Methods

Prior approval was obtained from the responsible research ethics committee in our institution (CAAE: 57064422.0.0000.5149) and from the airport security. Informed consent was obtained from participants. An experimental, non-randomized, controlled, cross-over study was carried out on August 12, 2022 at the Tancredo Neves International Airport in Belo Horizonte (MG), Brazil.

The airport walk-through metal detector (►**Fig. 1**) used was CEIA SMD600 PLUS (2021 model), device universally used in international airports. The experiment was performed in a device set for regular airport daily routine. Before starting, we made sure the device was working appropriately (►**Fig. 2**) by testing with metallic objects (belts and cellphones) and, at all times, the tests were performed under supervision of the responsible technician to ensure perfect functioning.

Two healthy volunteers, one male (172 cm in height) and one female (156 cm in height), who had no metal device in their body, were recruited. The volunteers walked in two different speeds each, one test at 2 km/hr and one at 6 km/hr. Initially, as a control, the volunteers walked across the metal detector without implants. Afterwards, we strapped orthopedic implants to the volunteers (►**Figs. 3** and **4**), using an adhesive tape, respecting as much as possible the true location of the implant in the body. All the tests were performed twice to ratify the results of the first pass.

Implants (►**Table 1**) used for fracture fixation and for joint replacement were assessed in different combinations (-unilateral/bilateral, left/right limbs or associated fractures) and always respecting the anatomic location (shoulder, arm, elbow, forearm, wrist, hip, thigh, knee, leg, and ankle). The implants for internal fixation were manufactured by Hexagon® (Itapira, SP, Brazil), Smith & Nephew (Memphis, TN, USA) and Tóride (Mogi Mirim, SP, Brazil), the external fixator by Baumer (Mogi Mirim, SP, Brazil), the hip prostheses by Baumer (Mogi Mirim, SP, Brazil) and Víncula (Rio Claro, SP,



Fig. 1 Airport walk-through metal detector (CEIA SMD600 PLUS, 2021 model).

Brazil), and the knee prostheses by Aesculap AG (Tuttlingen, Alemanha) and Baumer (Mogi Mirim, SP, Brazil).

Results

The results are presented in ►**Table 1**. From all osteosynthesis devices tested, the external fixator was the only one detected and its difference from the other devices was in its composition – it was made of aluminum alloy. All hip replacement implants tested were detected. Two knee replacement implants were tested, and both were made of cobalt-chromium-molybdenum, but with different specifications (►**Table 1**) and, only one of them was detected.

Discussion

Patients frequently ask about practical aspects of daily living and how certain types of surgery will affect their routine and for how long. Many of these questions still do not have a



Fig. 2 Control test to certify appropriate functioning of the arch detector using cellphone in the pocket.

definite answer based on scientific studies. An extremely common question is if retained orthopedic implants will trigger airport walk-through metal detectors. This has been a hot topic since September 11, 2001, when the World was shocked by airplanes terrorist attacks, after which airport security has increased to prevent such acts. The main concerns usually are the inconvenient of being body searched after triggering the metal detector, anxiety about being detained at the airport, and the possibility of delaying a travel or losing a flight. After 21 years of the terrorist attacks, to the best of our knowledge, there is still no official and universally accepted document that the patient can carry to prove the existence of an orthopedic metallic implant. Therefore, the standard procedure is further screening if an individual triggers a walk-through metal detector despite carrying a medical report issued by the orthopedic surgeon.[10] In Brazil, this is the first study to investigate orthopedic implant detection by airport walk-through metal detector. Additionally, all of the existing published researches were performed in arch detectors with more



Fig. 3 Total knee prosthesis strapped to the anterior surface of the right knee before testing.

than 5 years of manufacture, while in this study, a 2021 device, universally adopted in international airports, was used. It is noteworthy that this metal detector fully complies the current security level of the Tancredo Neves International Airport (Confins), under regulation of the ANAC (National Agency of Civil Aviation), and consequently of the International Civil Aviation Organization.

Our findings were completely unexpected, as none of the osteosynthesis sets were detected, except the external fixator device. No differences were observed between the two distinct transit speeds assessed. All tests were performed with implants on both right and left sides, to minimize the potential bias of the distance from the metal to the detector, and no differences were observed. All osteosynthesis sets tested were made of Stainless Steel NBR ISO 5832-1, titanium F-67 (distal radius plate) and titanium Ti-6Al-4V (Cephalomedullary femoral nail), except the external fixator which had aluminum alloy in its composition. The only Stainless Steel NBR ISO 5832-1 implant detected was the Thompson hip prosthesis, which suggests that implant mass concentration might increase detection. All other types of hip prosthesis tested also triggered the alarm. An interesting finding was the difference in detection between the two types of total knee prosthesis tested. Although both knee prostheses were made of cobalt-chromium-molybdenum, the detected one was manufactured in Brazil, while the other one in Germany. This difference among the knee prostheses tested raise a suspicion about the composition of the alloy, which could possibly have interfered on the arch detector triggering.

Considering all combinations of implants tested and presented in ►**Table 1**, our study corroborates with the findings of Chan et al.,[11] in which all patients with foot and ankle implants alone passed undetected. This study is consonant with several others of the literature, which state that total joint prostheses will be frequently detected, such as the hip prostheses in our study.[2,3,12] Kimura et al.[5] found that implant detection rate was higher during international flights, which might explain why all hip prostheses were detected in our study, since the device used was set for international standards, probably a more sensitive configuration.

Previous researches suggested some factors that might influence the probability of detection, such as implant mass, combinations, composition, location within the body, laterality, transit speed, detector model, sensitivity settings according to the security level of the airport, and tissue masking.[4,8,11] In our study, we were only capable to observe the probable interference of implant mass, density and material.

An interesting information about airport walk-through metal detectors is the fact that they record every individual that passes throw the arch even if there is no metal device in situ. This data is used in a randomizing internal software, which triggers a sound alarm to select a random passenger to be thoroughly body searched by the airport security. This sound alarm is different from the standard sound triggered when a metallic device is identified. This mechanism of selecting random passengers to be searched might be the reason why some patients with orthopedic implants report being body searched, which is a potential bias of previous research using retrospective information.[8]

The limitations of our study include the fact that none of the orthopedic implants were inside the volunteers tested and we still do not know if a bone-implant or soft tissues-implant interaction would affect detection. Previous researches tried to assess how the soft tissue envelope affect metal detection, however results were conflicting.[4,11] Another limitation is that we were not able to test all the available orthopedic implants. However, we judge that the most frequent situations were assessed. Our study also did not assess detection by portable wand metal detector, which is supposed to be more sensitive. Our findings were based on the current security level of the Confins International Airport. If, for some reason, the airport security level increases, the device's settings also change, and the detection sensitivity will improve. Although the tests were performed in a real environment of an international airport, following all the security requirements and under ANAC and ICAO rules, we cannot affirm that our findings will be faithfully reproducible if carried out at other airports, using a different detector device or a different security level.

However, some strengths deserve to be highlighted. This is the first study performed in Brazil aiming to clarify



Fig. 4 Bilateral retrograde femoral nail strapped to the lateral surface of the thighs before testing.

information regarding detection of orthopedic implants in an airport. We provided helpful information to patients, since the implants for internal fixation probably will not be detected in normal conditions. On the other hand, external fixators and prostheses will probably be detected. We also tried to minimize potential biases observed in similar studies. We tested different metal compositions in two healthy volunteers, under two different speeds to pass the detector arch, with implants attached in both sides (right and left) of the body, to minimize the bias of the distance from the detector to the implant.

Conclusions

In this study with ex-vivo orthopedic implants, we verified that osteosynthesis implants for internal fixation, composed by Stainless Steel ISO 5832-1 and titanium, were not detected by the airport walk-through metal detector and should not cause inconvenient to patients while travelling. However, external fixators and total joint prostheses will more frequently be detected.

It is important to highlight the limitations of our study and the need to further investigate this matter including in-vivo

Rev Bras Ortop Vol. 59 No. 5/2024 © 2023. The Author(s).

Table 1 Orthopedic implants sets and combinations tested

Implant	Male volunteer				Female volunteer				Alloy
	2 km/h		6 km/h		2 km/h		6 km/h		
	Test 1	Test 2	Test 1	Test 2	Test 1	Test 2	Test 1	Test 2	
Cellphone (control)	P	P	P	P	P	P	P	P	–
None (control)	N	N	N	N	N	N	N	N	–
Small fragment one-third tubular plate + 6 screws (unilateral)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Medial distal tibial locking compression plate + 6 screws (unilateral)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Medial and anterolateral distal tibial locking compression plate + 12 screws (unilateral)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Small fragment calcaneus plate + 6 screws (unilateral)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Intramedullary tibial nail + 4 locking screws (unilateral)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Intramedullary retrograde femoral nail + 3 locking screws (unilateral)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Intramedullary retrograde femoral nail + 3 locking screws (unilateral)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Bilateral intramedullary retrograde femoral nail + 6 locking screws + unilateral intramedullary tibial nail + 3 screws	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Cephalomedullary femoral nail + 1 sliding hip screw + 2 locking screws (unilateral)	N	N	N	N	N	N	N	N	Titanium Ti-6Al-4V (ASTM F1472)
Cephalomedullary femoral nail + 1 sliding hip screw + 2 locking screws (bilateral)	N	N	N	N	N	N	N	N	Titanium Ti-6Al-4V (ASTM F1472)
DCS with screws(unilateral)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
DHS with screws (unilateral)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Proximal humeral locking compression plate + 8 screws (unilateral)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Two 3.5 mm anchor screws on the shoulder (unilateral)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Intramedullary humeral nail + 3 locking screws (unilateral)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1

Table 1 (*Continued*)

Implant	Male volunteer				Female volunteer				Alloy
	2 km/h		6 km/h		2 km/h		6 km/h		
	Test 1	Test 2	Test 1	Test 2	Test 1	Test 2	Test 1	Test 2	
Clavicle locking compression plate + 6 screws (unilateral)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Clavicle locking compression plate + 6 screws (bilateral)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
12 holes DCP + 4 screws (unilateral humerus)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Two 3.5 mm cannulated screws + cerclage wire (1.0 mm) (unilateral knee)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Two 3.5 mm cannulated screws + cerclage wire (1.0 mm) (bilateral knee)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Two 7.0 mm cannulated screws + 1 cerclage wire (1.0 mm) (unilateral knee)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Two 7.0 mm cannulated screws + 1 cerclage wire (1.0 mm) (bilateral knee)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Distal radius locking compression plate + 11 screws (unilateral)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Distal radius locking compression plate + 11 screws (bilateral)	N	N	N	N	N	N	N	N	Stainless steel NBR ISO 5832-1
Distal radius locking compression plate + 11 screws (unilateral)	N	N	N	N	N	N	N	N	Titanium (ASTM F-67)
Distal radius locking compression plate + 11 screws (bilateral)	N	N	N	N	N	N	N	N	Titanium (ASTM F-67)
Cemented primary total knee prosthesis (unilateral)	P	P	P	P	P	P	P	P	Cobalt-chromium-molybdenum (ISO 5832-4)
Cemented primary total knee prosthesis (unilateral)	N	N	N	N	N	N	N	N	Cobalt-chromium-molybdenum (ISO 5832-4)
Thompson hip prosthesis (unilateral)	P	P	P	P	P	P	P	P	Stainless steel NBR ISO 5832-1

(*Continued*)

Rev Bras Ortop Vol. 59 No. 5/2024 © 2023. The Author(s).

Table 1 (*Continued*)

Implant	Male volunteer				Female volunteer				Alloy
	2 km/h		6 km/h		2 km/h		6 km/h		
	Test 1	Test 2	Test 1	Test 2	Test 1	Test 2	Test 1	Test 2	
Uncemented acetabulum + uncemented femoral stem (unilateral)	P	P	P	P	P	P	P	P	Titanium (ASTM F-67) (acetabulum) + Titanium Ti-6A-4V (ASTM F-136) and titanium porous coating (ASTM F-1580) (femoral component)
Uncemented acetabulum + cemented primary femoral stem (unilateral)	P	P	P	P	P	P	P	P	Titanium (ASTM F-67) (acetabulum) + Stainless steel (NBR ISO 5832-9/ ASTM F-1586)
External fixator (4 Schanz screws + 2 bars + 8 self-holding clamps)	P	P	P	P	P	P	P	P	Stainless steel (ISO 5832-1, ASTM F-138) and aluminum alloy

Abbreviations: ASTM, American Society for Testing and Materials; N, negative; NBR ISO, Norma Brasileira Regulamentadora - International Organization for Standardization; P, positive.

orthopedic implants and testing prostheses with different sizes, weights and materials.

Financial Support
The authors declare that this study did not receive any financial support from public funding, commercial or non-profit organization.

Conflict of Interests
IGNR declares no conflicts of interest. BMG declares no conflicts of interest. SHFZ declares no conflicts of interest. MAPA is speaker and lecturer of Zimmer-Biomet and AO Foundation. RESP is speaker and lecturer of Zimmer Biomet, Smith & Nephew and AO Foundation.

Acknowledgments
We would like to thank the Tancredo Neves International Airport of Belo Horizonte, the company BH Airport, the ANAC and all the security agencies of the airport, and specially to Mr. Wesley Dias Santos for the availability and for the help to perform the tests in the airport premises, providing all the technical information to this study possible. Additionally, we would like to thank AMGS Comércio e Representações Ltda and Mr. Ricardo Julião for providing the orthopedic implants and the technical information about the implants to be tested in this study and, Ms. Elen Rocha for helping to assemble the kits to be tested.

References

1 Dines JS, Elkousy H, Edwards TB, Gartsman GM, Dines DM. Effect of total shoulder replacements on airport security screening in the post-9/11 era. J Shoulder Elbow Surg 2007;16(04):434–437
2 Obremskey WT, Austin T, Crosby C, et al. Detection of orthopaedic implants by airport metal detectors. J Orthop Trauma 2007;21 (02):129–132
3 Ramirez MA, Rodriguez EK, Zurakowski D, Richardson LC. Detection of orthopaedic implants in vivo by enhanced-sensitivity, walk-through metal detectors. J Bone Joint Surg Am 2007;89 (04):742–746
4 Kuczmarski AS, Harris AP, Gil JA, Owens BD. Sensitivity of Airport Metal Detectors to Orthopaedic Implants. JBJS Rev 2018;6(07):e7
5 Kimura A, Jinno T, Tsukada S, Matsubara M, Koga H. Detection of total hip prostheses at airport security checkpoints. J Orthop Sci 2020;25(02):255–260
6 Issa K, Pierce TP, Gwam C, Festa A, Scillia AJ, Mont MA. Detection of total hip arthroplasties at airport security checkpoints - how do updated security measures affect patients? Hip Int 2018;28(02): 122–124
7 Issa K, Pierce TP, Gwam C, et al. Detection of Total Knee Arthroplasties at Airport Security Checkpoints: How Do Updated Security Measures Affect Patients? J Knee Surg 2017;30(06):532–534
8 Johnson AJ, Naziri Q, Hooper HA, Mont MA. Detection of total hip prostheses at airport security checkpoints: how has heightened security affected patients? J Bone Joint Surg Am 2012;94(07):e44
9 Abbassian A, Datla B, Brooks RA. Detection of orthopaedic implants by airport metal detectors. Ann R Coll Surg Engl 2007; 89(03):285–287
10 TSA. What are the procedures if I have an internal or external medical device, such as a pacemaker or metal implant? [accessed 13th September, 2022]. Available from: https://www.tsa.gov/travel/frequently-asked-questions/what-are-procedures-if-i-have-internal-or-external-medical-device
11 Chan JY, Mani SB, Williams PN, et al. Detection of In Vivo Foot and Ankle Implants by Walkthrough Metal Detectors. Foot Ankle Int 2014;35(08):789–795
12 Naziri Q, Johnson AJ, Hooper HA, Sana SH, Mont MA. Detection of total knee prostheses at airport security checkpoints. J Arthroplasty 2012;27(06):1228–1233